

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071918\
 Data File : VV006645.D
 Acq On : 19 Jul 2018 16:21
 Operator : SY/MD
 Sample : J3983-19
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB199

Quant Time: Jul 20 05:15:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 20 05:09:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	254250	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	239872	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	111978	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	95246	5.34	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.80%
7) Chloroethane-d5	1.58	69	76699	6.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	127.40%
11) 1,1-Dichloroethene-d2	2.13	63	123254	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	3.96	46	225356	58.15	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.30%
24) Chloroform-d	4.40	84	163513	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.00%
26) 1,2-Dichloroethane-d4	5.09	65	80514	5.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.60%
32) Benzene-d6	5.10	84	345460	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.00%
36) 1,2-Dichloropropane-d6	6.12	67	112923	5.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.80%
41) Toluene-d8	7.36	98	310913	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
43) trans-1,3-Dichloropropene-	7.67	79	34911	5.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.80%
46) 2-Hexanone-d5	8.14	63	211767	61.96	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.92%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	81830	5.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	113647	5.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.80%

Target Compounds

					Ovalue
13) Acetone	2.21	43	84567	28.45	ug/L 91
14) Carbon disulfide	2.32	76	19931	0.33	ug/L 96
15) Methyl Acetate	2.47	43	749006	85.89	ug/L 99
18) trans-1,2-Dichloroethene	2.79	96	55987	2.70	ug/L 99
21) 2-Butanone	4.04	43	79319	15.93	ug/L 99
47) Tetrachloroethene	8.02	164	4206	0.25	ug/L 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV071918\
Data File : VV006645.D
Acq On : 19 Jul 2018 16:21
Operator : SY/MD
Sample : J3983-19
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB199

Quant Time: Jul 20 05:15:35 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 20 05:09:32 2018
Response via : Initial Calibration

