

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW071922\
 Data File : VW026830.D
 Acq On : 19 Jul 2022 16:43
 Operator : SY/MD
 Sample : N3754-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBHP7

Quant Time: Jul 20 05:51:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 20 05:46:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	209274	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	220530	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	86652	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	68467	3.586	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.800%
7) Chloroethane-d5	1.561	69	67439	4.088	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.800%
11) 1,1-Dichloroethene-d2	2.101	63	93692	2.556	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.200%#
20) 2-Butanone-d5	3.883	46	179415	60.212	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.420%
24) Chloroform-d	4.342	84	160038	4.794	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.800%
26) 1,2-Dichloroethane-d4	5.031	65	73993	5.074	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.400%
32) Benzene-d6	5.047	84	266413	3.976	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.600%
36) 1,2-Dichloropropane-d6	6.066	67	95819	4.769	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	7.317	98	186357	3.133	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	62.600%#
43) trans-1,3-Dichloroprop...	7.625	79	24374	3.708	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.200%
46) 2-Hexanone-d5	8.088	63	131222	46.106	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.220%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	74927	4.787	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	76386	5.059	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071922\
Data File : VV026830.D
Acq On : 19 Jul 2022 16:43
Operator : SY/MD
Sample : N3754-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBHP7

Quant Time: Jul 20 05:51:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 20 05:46:20 2022
Response via : Initial Calibration

