

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW071922\
 Data File : VW026837.D
 Acq On : 19 Jul 2022 19:58
 Operator : SY/MD
 Sample : N3754-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBHQ3

Quant Time: Jul 20 05:53:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 20 05:46:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	217946	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	227576	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	91956	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	66445	3.342	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.800%
7) Chloroethane-d5	1.564	69	67543	3.932	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.600%
11) 1,1-Dichloroethene-d2	2.101	63	90809	2.379	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	47.600%#
20) 2-Butanone-d5	3.892	46	191328	61.655	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.320%
24) Chloroform-d	4.346	84	162004	4.660	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
26) 1,2-Dichloroethane-d4	5.031	65	76352	5.027	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.600%
32) Benzene-d6	5.047	84	264764	3.829	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.600%
36) 1,2-Dichloropropane-d6	6.069	67	95569	4.609	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.200%
41) Toluene-d8	7.317	98	184255	3.002	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	60.000%#
43) trans-1,3-Dichloroprop...	7.625	79	24747	3.648	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.000%
46) 2-Hexanone-d5	8.088	63	139040	47.341	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.680%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	79166	4.901	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	77786	4.855	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071922\
Data File : VV026837.D
Acq On : 19 Jul 2022 19:58
Operator : SY/MD
Sample : N3754-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBHQ3

Quant Time: Jul 20 05:53:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 20 05:46:20 2022
Response via : Initial Calibration

