

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW071922\  
 Data File : VW026828.D  
 Acq On : 19 Jul 2022 15:53  
 Operator : SY/MD  
 Sample : N3754-04  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 GBHP5

Quant Time: Jul 20 05:50:52 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR071122WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed Jul 20 05:46:20 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.616  | 114   | 230920   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.850  | 117   | 238610   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 94242    | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.301  | 65    | 72745    | 3.453    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 69.000%  |
| 7) Chloroethane-d5            | 1.561  | 69    | 73247    | 4.024    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 80.400%  |
| 11) 1,1-Dichloroethene-d2     | 2.102  | 63    | 100759   | 2.491    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 49.800%# |
| 20) 2-Butanone-d5             | 3.899  | 46    | 199815   | 60.772   | ug/L  | 0.01     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 121.540% |
| 24) Chloroform-d              | 4.346  | 84    | 167337   | 4.543    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 90.800%  |
| 26) 1,2-Dichloroethane-d4     | 5.031  | 65    | 80141    | 4.980    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 99.600%  |
| 32) Benzene-d6                | 5.047  | 84    | 285540   | 3.938    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 78.800%  |
| 36) 1,2-Dichloropropane-d6    | 6.066  | 67    | 103005   | 4.738    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 94.800%  |
| 41) Toluene-d8                | 7.314  | 98    | 201300   | 3.128    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 62.600%# |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79    | 27136    | 3.815    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 76.400%  |
| 46) 2-Hexanone-d5             | 8.088  | 63    | 147647   | 47.947   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 95.900%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 81469    | 4.811    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 96.200%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152   | 82396    | 5.018    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 100.400% |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV071922\  
Data File : VV026828.D  
Acq On : 19 Jul 2022 15:53  
Operator : SY/MD  
Sample : N3754-04  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
GBHP5

Quant Time: Jul 20 05:50:52 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR071122WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Wed Jul 20 05:46:20 2022  
Response via : Initial Calibration

