

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071923\
 Data File : VV031722.D
 Acq On : 19 Jul 2023 18:48
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5298

Quant Time: Jul 20 05:13:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM071923.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 20 05:08:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	5.542	114	4465	0.500 ug/L	0.00
5) Chlorobenzene-d5	8.800	117	4305	0.500 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2193	0.500 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.280	65	2295	0.646 ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 130.000%		
4) 1,2-Dichloroethane-d4	4.970	65	1838	0.631 ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 126.000%		
7) 1,2-Dichloropropane-d6	5.996	67	2010	0.576 ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 116.000%		
8) Toluene-d8	7.254	98	3074	0.535 ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 108.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1596	0.586 ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 118.000%		
Target Compounds					
					Qvalue
3) Vinyl chloride	1.280	62	3500	0.624 ug/L	99
6) Trichloroethene	5.852	95	1800	0.528 ug/L	99
9) 1,2-Dibromoethane	8.290	107	1368	0.552 ug/L	99
12) 1,2,3-Trichloropropane	10.214	75	1191	0.546 ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	210	0.504 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071923\
Data File : VV031722.D
Acq On : 19 Jul 2023 18:48
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5298

Quant Time: Jul 20 05:13:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM071923.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 20 05:08:22 2023
Response via : Initial Calibration

