

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071923\
 Data File : VV031715.D
 Acq On : 19 Jul 2023 15:02
 Operator : SY/MD
 Sample : VSTD00135
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001235

Quant Time: Jul 20 05:00:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM071923.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 20 04:59:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	4886	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	4750	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2381	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	4485	1.038	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.946	65	3519	1.103	ug/L	-0.02
7) 1,2-Dichloropropane-d6	5.995	67	4165	1.012	ug/L	0.00
8) Toluene-d8	7.253	98	6822	0.845	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3105	1.121	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	7008	0.967	ug/L	99
6) Trichloroethene	5.828	95	4102	0.963	ug/L	86
9) 1,2-Dibromoethane	8.290	107	2891	1.146	ug/L #	96
12) 1,2,3-Trichloropropane	10.197	75	2471	1.010	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	448	1.191	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071923\
Data File : VV031715.D
Acq On : 19 Jul 2023 15:02
Operator : SY/MD
Sample : VSTD00135
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001235

Quant Time: Jul 20 05:00:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM071923.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 20 04:59:20 2023
Response via : Initial Calibration

