

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072018\
 Data File : VV006668.D
 Acq On : 20 Jul 2018 15:32
 Operator : SY/MD
 Sample : J4078-01
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA0

Quant Time: Jul 26 03:02:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 26 02:08:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	228383	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	212026	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	92682	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	89301	5.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.40%
7) Chloroethane-d5	1.58	69	66249	6.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	122.60%
11) 1,1-Dichloroethene-d2	2.13	63	122639	3.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.60%
20) 2-Butanone-d5	3.95	46	178734	51.35	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.70%
24) Chloroform-d	4.41	84	148250	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.00%
26) 1,2-Dichloroethane-d4	5.09	65	70920	5.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.40%
32) Benzene-d6	5.10	84	319382	5.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.00%
36) 1,2-Dichloropropane-d6	6.12	67	99222	5.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.00%
41) Toluene-d8	7.36	98	279860	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
43) trans-1,3-Dichloropropene-	7.67	79	28158	4.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.00%
46) 2-Hexanone-d5	8.14	63	157605	52.17	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.34%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	68979	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	97270	5.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.60%

Target Compounds					Ovalue
13) Acetone	2.19	43	7019	2.63 ug/L	94
14) Carbon disulfide	2.32	76	19068	0.35 ug/L	97
15) Methyl Acetate	2.46	43	3136	0.40 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072018\
Data File : VV006668.D
Acq On : 20 Jul 2018 15:32
Operator : SY/MD
Sample : J4078-01
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA0

Quant Time: Jul 26 03:02:54 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 26 02:08:26 2018
Response via : Initial Calibration

