

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072018\
 Data File : VV006669.D
 Acq On : 20 Jul 2018 16:00
 Operator : SY/MD
 Sample : J3983-19DL 5X
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB199DL

Quant Time: Jul 26 03:05:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 26 02:08:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	223057	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	206854	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	87952	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	83770	5.35	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	107.00%
7) Chloroethane-d5	1.58	69	60241	5.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.00%
11) 1,1-Dichloroethene-d2	2.13	63	116649	3.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.60%
20) 2-Butanone-d5	3.95	46	139330	40.98	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.96%
24) Chloroform-d	4.41	84	140582	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.80%
26) 1,2-Dichloroethane-d4	5.09	65	66867	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.40%
32) Benzene-d6	5.10	84	302476	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.80%
36) 1,2-Dichloropropane-d6	6.12	67	93649	5.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.40%
41) Toluene-d8	7.36	98	263997	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
43) trans-1,3-Dichloropropene-	7.67	79	25758	4.36	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.20%
46) 2-Hexanone-d5	8.14	63	143119	48.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.12%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	64262	5.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	89817	5.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.40%

Target Compounds					Ovalue
13) Acetone	2.19	43	7682	2.95 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072018\
Data File : VV006669.D
Acq On : 20 Jul 2018 16:00
Operator : SY/MD
Sample : J3983-19DL 5X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB199DL

Quant Time: Jul 26 03:05:03 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 26 02:08:26 2018
Response via : Initial Calibration

