

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072018\
 Data File : VV006670.D
 Acq On : 20 Jul 2018 16:27
 Operator : SY/MD
 Sample : J4078-02
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA3

Quant Time: Jul 26 03:07:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 26 02:08:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	234323	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	220877	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	97856	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	82851	5.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.80%
7) Chloroethane-d5	1.58	69	58042	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.60%
11) 1,1-Dichloroethene-d2	2.13	63	109769	3.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.40%
20) 2-Butanone-d5	3.95	46	156364	43.78	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.56%
24) Chloroform-d	4.41	84	137373	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
26) 1,2-Dichloroethane-d4	5.09	65	65750	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
32) Benzene-d6	5.10	84	294849	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	6.12	67	92421	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.36	98	268598	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	7.67	79	29236	4.63	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.60%
46) 2-Hexanone-d5	8.14	63	153582	48.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.60%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	65247	4.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	92388	5.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.00%

Target Compounds

					Ovalue
13) Acetone	2.19	43	23365	8.53 ug/L	99
14) Carbon disulfide	2.32	76	14496	0.26 ug/L	96
15) Methyl Acetate	2.46	43	26574	3.31 ug/L	100
16) Methylene chloride	2.54	84	6823	0.31 ug/L	90
18) trans-1,2-Dichloroethene	2.79	96	9249	0.48 ug/L	93
21) 2-Butanone	4.04	43	15393	3.36 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072018\
Data File : VV006670.D
Acq On : 20 Jul 2018 16:27
Operator : SY/MD
Sample : J4078-02
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA3

Quant Time: Jul 26 03:07:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 26 02:08:26 2018
Response via : Initial Calibration

