

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072018\
 Data File : VV006671.D
 Acq On : 20 Jul 2018 16:54
 Operator : SY/MD
 Sample : J3983-20
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB1A1

Quant Time: Jul 26 03:43:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 26 02:08:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	235487	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	215198	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	96345	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	82987	5.02	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.40%
7) Chloroethane-d5	1.57	69	58606	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.20%
11) 1,1-Dichloroethene-d2	2.13	63	111457	3.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.20%
20) 2-Butanone-d5	3.95	46	169894	47.33	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.66%
24) Chloroform-d	4.40	84	137525	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	5.09	65	66015	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
32) Benzene-d6	5.10	84	293061	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.40%
36) 1,2-Dichloropropane-d6	6.12	67	91758	5.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.20%
41) Toluene-d8	7.36	98	255842	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.20%
43) trans-1,3-Dichloropropene-	7.67	79	24818	4.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.80%
46) 2-Hexanone-d5	8.14	63	148015	48.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.54%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	64604	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	90735	5.23	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.60%

Target Compounds

					Ovalue
13) Acetone	2.19	43	3296	1.20 ug/L	97
14) Carbon disulfide	2.32	76	18734	0.33 ug/L	96
18) trans-1,2-Dichloroethene	2.79	96	335960	17.48 ug/L	98
22) cis-1,2-Dichloroethene	3.96	96	572966	28.48 ug/L	100
34) Trichloroethene	5.96	95	188902	9.86 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072018\
Data File : VV006671.D
Acq On : 20 Jul 2018 16:54
Operator : SY/MD
Sample : J3983-20
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB1A1

Quant Time: Jul 26 03:43:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 26 02:08:26 2018
Response via : Initial Calibration

