

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072018\
 Data File : VV006657.D
 Acq On : 20 Jul 2018 10:01
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00571

Quant Time: Jul 26 02:03:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun Jul 22 23:06:02 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.475	0.532	-12.0	104	0.00
3 T	Chloromethane	0.402	0.434	-8.0	102	0.00
4 S	Vinyl Chloride-d3	0.351	0.347	1.1	94	0.00
5 T	Vinyl chloride	0.519	0.554	-6.7	101	0.00
6 T	Bromomethane	0.147	0.147	0.0	98	0.00
7 S	Chloroethane-d5	0.237	0.263	-11.0	100	0.00
8 T	Chloroethane	0.298	0.326	-9.4	105	0.00
9 T	Trichlorofluoromethane	0.623	0.701	-12.5	106	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.391	0.442	-13.0	108	0.00
11 S	1,1-Dichloroethene-d2	0.675	0.704	-4.3	101	0.00
12 T	1,1-Dichloroethene	0.370	0.408	-10.3	106	0.00
13 T	Acetone	0.058	0.059	-1.7	97	0.00
14 T	Carbon disulfide	1.200	1.321	-10.1	106	0.00
15 T	Methyl Acetate	0.171	0.172	-0.6	97	0.00
16 T	Methylene chloride	0.466	0.452	3.0	105	0.00
17 T	Methyl tert-butyl Ether	0.914	0.971	-6.2	101	0.00
18 T	trans-1,2-Dichloroethene	0.408	0.447	-9.6	104	0.00
19 T	1,1-Dichloroethane	0.720	0.819	-13.7	108	0.00
20 S	2-Butanone-d5	0.076	0.051	32.9#	63	0.00
21 T	2-Butanone	0.098	0.102	-4.1	98	0.00
22 T	cis-1,2-Dichloroethene	0.427	0.474	-11.0	105	0.00
23 T	Bromochloromethane	0.174	0.191	-9.8	104	0.00
24 S	Chloroform-d	0.596	0.617	-3.5	98	0.00
25 T	Chloroform	0.694	0.792	-14.1	108	0.00
26 S	1,2-Dichloroethane-d4	0.279	0.286	-2.5	97	0.00
27 T	1,2-Dichloroethane	0.406	0.447	-10.1	104	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
29 T	1,1,1-Trichloroethane	0.625	0.705	-12.8	107	0.00
30 T	Cyclohexane	0.699	0.784	-12.2	105	0.00
31 T	Carbon tetrachloride	0.535	0.601	-12.3	106	0.00
32 S	Benzene-d6	1.357	1.435	-5.7	99	0.00
33 T	Benzene	1.759	1.992	-13.2	106	0.00
34 T	Trichloroethene	0.445	0.487	-9.4	106	0.00
35 T	Methylcyclohexane	0.743	0.844	-13.6	106	0.00
36 S	1,2-Dichloropropane-d6	0.421	0.447	-6.2	101	0.00
37 T	1,2-Dichloropropane	0.449	0.506	-12.7	105	0.00
38 T	Bromodichloromethane	0.498	0.548	-10.0	104	0.00
39 T	cis-1,3-Dichloropropene	0.590	0.660	-11.9	104	0.00
40 T	4-Methyl-2-pentanone	0.243	0.253	-4.1	95	0.00
41 S	Toluene-d8	1.263	1.355	-7.3	100	0.00
42 T	Toluene	1.818	2.077	-14.2	106	0.00
43 S	trans-1,3-Dichloropropene-d	0.143	0.150	-4.9	99	0.00
44 T	trans-1,3-Dichloropropene	0.459	0.502	-9.4	101	0.00
45 T	1,1,2-Trichloroethane	0.302	0.325	-7.6	102	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072018\
 Data File : VV006657.D
 Acq On : 20 Jul 2018 10:01
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00571

Quant Time: Jul 26 02:03:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun Jul 22 23:06:02 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	0.071	0.067	5.6	86	0.00
47 T	Tetrachloroethene	0.357	0.403	-12.9	107	0.00
48 T	2-Hexanone	0.170	0.181	-6.5	96	0.00
49 T	Dibromochloromethane	0.320	0.357	-11.6	104	0.00
50 T	1,2-Dibromoethane	0.271	0.288	-6.3	99	0.00
51 T	Chlorobenzene	1.181	1.325	-12.2	106	0.00
52 T	Ethylbenzene	1.952	2.240	-14.8	106	0.00
53 T	m,p-xylene	0.742	0.869	-17.1	108	0.00
54 T	o-xylene	0.724	0.840	-16.0	107	0.00
55 T	Styrene	1.212	1.427	-17.7	106	0.00
56 T	Isopropylbenzene	1.899	2.211	-16.4	106	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.306	0.312	-2.0	96	0.00
58 T	1,1,2,2-Tetrachloroethane	0.362	0.392	-8.3	101	0.00
59	1,2,3-Trichloropropane	0.265	0.278	-4.9	100	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
61 T	Bromoform	0.346	0.362	-4.6	101	0.00
62 T	1,3-Dichlorobenzene	1.821	2.030	-11.5	106	0.00
63 T	1,4-Dichlorobenzene	1.850	2.048	-10.7	106	0.00
64 S	1,2-Dichlorobenzene-d4	0.900	0.951	-5.7	102	0.00
65 T	1,2-Dichlorobenzene	1.746	1.941	-11.2	105	0.00
66 T	1,2-Dibromo-3-chloropropane	0.104	0.108	-3.8	97	0.00
67	1,3,5-Trichlorobenzene	1.408	1.602	-13.8	108	0.00
68 T	1,2,4-trichlorobenzene	1.117	1.240	-11.0	104	0.00
69	Naphthalene	1.933	1.984	-2.6	95	0.00
70 T	1,2,3-Trichlorobenzene	1.052	1.158	-10.1	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0