


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072021\
Data File : VV021691.D
Acq On : 20 Jul 2021 16:36
Operator : SY/MD
Sample : M3034-02ME
Misc : 5.06g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AT4ME

Manual Integrations
APPROVED

MMDadoda
7/23/2021 12:58:08 PM

Quant Time: Jul 21 01:06:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071921WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 21 01:01:54 2021
Response via : Initial Calibration



Quantitation Report (Qedit)


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072021\
 Data File : VV021691.D
 Acq On : 20 Jul 2021 16:36
 Operator : SY/MD
 Sample : M3034-02ME
 Misc : 5.06g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COAT4ME

Manual Integrations
 APPROVED

MMDadoda
 7/23/2021 12:58:08 PM

Quant Time: Jul 21 01:06:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 21 01:01:54 2021
 Response via : Initial Calibration



(7) Chloroethane-d5 (S)

1.532min (-0.032) 10.23 ug/L

response 25607

Ion	Exp%	Act%
69.00	100.00	100.00
71.00	32.20	47.76#
51.00	23.40	17.82
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072021\
 Data File : VV021691.D
 Acq On : 20 Jul 2021 16:36
 Operator : SY/MD
 Sample : M3034-02ME
 Misc : 5.06g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COAT4ME

Manual Integrations
 APPROVED

MMDadoda
 7/23/2021 12:58:08 PM

Quant Time: Jul 21 01:06:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 21 01:01:54 2021
 Response via : Initial Calibration



TIC: VV021691.D\data.ms

(7) Chloroethane-d5 (S)

1.532min (-0.032) 14.42 ug/L m

response 36096

Ion	Exp%	Act%
69.00	100.00	100.00
71.00	32.20	33.88
51.00	23.40	12.64#
0.00	0.00	0.00

7MD
7/23/21

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072021\
 Data File : VV021691.D
 Acq On : 20 Jul 2021 16:36
 Operator : SY/MD
 Sample : M3034-02ME
 Misc : 5.06g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COAT4ME

Manual Integrations
 APPROVED

MMDadoda
 7/23/2021 12:58:08 PM

Quant Time: Jul 21 01:06:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 21 01:01:54 2021
 Response via : Initial Calibration



TIC: VV021691.D\data.ms

(47) 2-Hexanone-d5 (S)

8.101min (+ 0.013) 50.99 ug/L

response 102396

Ion	Exp%	Act%
63.00	100.00	100.00
46.00	163.60	164.84
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\V072021\
 Data File : V0721691.D
 Acq On : 20 Jul 2021 16:36
 Operator : SY/MD
 Sample : M3034-02ME
 Misc : 5.06g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AT4ME

Manual Integrations
 APPROVED

MMDadoda
 7/23/2021 12:58:08 PM

Quant Time: Jul 21 01:06:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 21 01:01:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	511348	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	483178	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	237582	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	191270	52.875	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 105.760%			
7) Chloroethane-d5	1.532	69	36096m	14.422	ug/L	-0.03
Spiked Amount 50.000	Range 70 - 130		Recovery = 28.840%#			
11) 1,1-Dichloroethene-d2	2.066	63	238839	39.622	ug/L	-0.04
Spiked Amount 50.000	Range 60 - 125		Recovery = 79.240%			
21) 2-Butanone-d5	3.892	46	280845	100.728	ug/L	0.02
Spiked Amount 100.000	Range 40 - 130		Recovery = 100.730%			
24) Chloroform-d	4.346	84	351389	51.156	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 102.320%			
26) 1,2-Dichloroethane-d4	5.027	65	220988	54.316	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 108.640%			
32) Benzene-d6	5.040	84	751613	52.852	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 105.700%			
36) 1,2-Dichloropropane-d6	6.069	67	237341	53.378	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 106.760%			
41) Toluene-d8	7.313	98	693548	52.787	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 105.580%			
43) trans-1,3-Dichloroprop...	7.622	79	114942	51.857	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 103.720%			
47) 2-Hexanone-d5	8.120	63	198722m	98.959	ug/L	0.03
Spiked Amount 100.000	Range 45 - 130		Recovery = 98.960%			
56) 1,1,2,2-Tetrachloroeth...	10.223	84	300956	51.241	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 102.480%			
66) 1,2-Dichlorobenzene-d4	11.631	152	271863	55.352	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 110.700%			
Target Compounds						
16) Methylene chloride	2.490	84	4732	1.215	ug/L	98
25) Chloroform	4.368	83	4347	0.693	ug/L #	36

(#) = qualifier out of range (m) = manual integration (+) = signals summed

7/23/21