

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072022\
 Data File : VW026850.D
 Acq On : 20 Jul 2022 15:32
 Operator : SY/MD
 Sample : N3802-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBHW7

Quant Time: Jul 21 01:50:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jul 21 01:47:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	224272	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	231104	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	95669	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	65122	3.183	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.600%
7) Chloroethane-d5	1.564	69	66792	3.778	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.600%
11) 1,1-Dichloroethene-d2	2.105	63	91367	2.326	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	46.600%#
20) 2-Butanone-d5	3.883	46	180562	56.545	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.080%
24) Chloroform-d	4.346	84	165949	4.639	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.030	65	74995	4.798	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
32) Benzene-d6	5.047	84	268995	3.831	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.600%
36) 1,2-Dichloropropane-d6	6.069	67	97665	4.638	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.800%
41) Toluene-d8	7.313	98	188419	3.023	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	60.400%#
43) trans-1,3-Dichloroprop...	7.625	79	24753	3.593	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.800%
46) 2-Hexanone-d5	8.088	63	139250	46.688	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.380%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	79927	4.873	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	81327	4.879	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.600%
Target Compounds						
					Qvalue	
10) 1,1,2-Trichloro-1,2,2-...	2.114	101	43182	2.129	ug/L	99
22) cis-1,2-Dichloroethene	3.915	96	12709	0.697	ug/L	90
34) Trichloroethene	5.911	95	145019	7.637	ug/L	99
47) Tetrachloroethene	7.972	164	17566	1.179	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072022\
Data File : VV026850.D
Acq On : 20 Jul 2022 15:32
Operator : SY/MD
Sample : N3802-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBHW7

Quant Time: Jul 21 01:50:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jul 21 01:47:16 2022
Response via : Initial Calibration

