

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072022\
 Data File : VW026854.D
 Acq On : 20 Jul 2022 17:08
 Operator : SY/MD
 Sample : N3802-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBHX0

Quant Time: Jul 21 01:51:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jul 21 01:47:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	216863	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	222701	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	95564	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	68547	3.465	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.200%
7) Chloroethane-d5	1.564	69	68924	4.032	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.600%
11) 1,1-Dichloroethene-d2	2.104	63	93595	2.464	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	49.200%#
20) 2-Butanone-d5	3.889	46	200766	65.020	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	130.040%#
24) Chloroform-d	4.352	84	173311	5.010	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.034	65	78931	5.223	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.050	84	276459	4.085	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.800%
36) 1,2-Dichloropropane-d6	6.072	67	101146	4.985	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.800%
41) Toluene-d8	7.317	98	194867	3.244	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	64.800%#
43) trans-1,3-Dichloroprop...	7.628	79	26361	3.971	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.400%
46) 2-Hexanone-d5	8.091	63	143469	49.918	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.840%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	79998	5.061	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	85190	5.116	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072022\
Data File : VV026854.D
Acq On : 20 Jul 2022 17:08
Operator : SY/MD
Sample : N3802-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBHX0

Quant Time: Jul 21 01:51:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jul 21 01:47:16 2022
Response via : Initial Calibration

