

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072023\
 Data File : VV031724.D
 Acq On : 20 Jul 2023 10:33
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5299

Quant Time: Jul 21 01:23:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM071923.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 20 05:08:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	5.542	114	4636	0.500 ug/L	0.00
5) Chlorobenzene-d5	8.800	117	4402	0.500 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2163	0.500 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.280	65	1942	0.526 ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 106.000%		
4) 1,2-Dichloroethane-d4	4.970	65	1699	0.562 ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 112.000%		
7) 1,2-Dichloropropane-d6	5.996	67	1930	0.541 ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 108.000%		
8) Toluene-d8	7.254	98	2984	0.508 ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 102.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1483	0.533 ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 106.000%		
Target Compounds					
					Qvalue
3) Vinyl chloride	1.280	62	3114	0.534 ug/L	99
6) Trichloroethene	5.852	95	1868	0.536 ug/L	98
9) 1,2-Dibromoethane	8.290	107	1391	0.549 ug/L	99
12) 1,2,3-Trichloropropane	10.214	75	1136	0.528 ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	203	0.494 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072023\
Data File : VV031724.D
Acq On : 20 Jul 2023 10:33
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5299

Quant Time: Jul 21 01:23:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM071923.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 20 05:08:22 2023
Response via : Initial Calibration

