

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072118\
 Data File : VV006683.D
 Acq On : 21 Jul 2018 13:42
 Operator : SY/MD
 Sample : J4078-09
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AB6

Quant Time: Jul 22 23:17:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun Jul 22 23:06:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	219001	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	203339	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	83168	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	79341	5.16	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.20%
7) Chloroethane-d5	1.58	69	60744	5.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.20%
11) 1,1-Dichloroethene-d2	2.13	63	112283	3.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.00%
20) 2-Butanone-d5	3.96	46	71318	21.37	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	42.74%
24) Chloroform-d	4.41	84	134212	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	5.09	65	66545	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.00%
32) Benzene-d6	5.10	84	294207	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.60%
36) 1,2-Dichloropropane-d6	6.12	67	89708	5.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.80%
41) Toluene-d8	7.37	98	252667	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
43) trans-1,3-Dichloropropene-	7.67	79	23618	4.06	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.20%
46) 2-Hexanone-d5	8.14	63	127067	43.86	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.72%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	61277	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	84790	5.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.40%

Target Compounds					Ovalue
13) Acetone	2.19	43	7876	3.08 ug/L	98
21) 2-Butanone	4.02	43	18390	4.29 ug/L	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072118\
Data File : VV006683.D
Acq On : 21 Jul 2018 13:42
Operator : SY/MD
Sample : J4078-09
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AB6

Quant Time: Jul 22 23:17:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR071918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sun Jul 22 23:06:02 2018
Response via : Initial Calibration

