

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072120\  
 Data File : VV017567.D  
 Acq On : 21 Jul 2020 16:33  
 Operator : SY/MD  
 Sample : VV0721WBL01  
 Misc : 5.0mL/MSVOA V/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK64

Quant Time: Jul 22 03:35:38 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVLM072120WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Wed Jul 22 03:23:09 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.64  | 114  | 329679   | 50.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 8.87  | 117  | 313572   | 50.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.27 | 152  | 153496   | 50.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |         |       |          |          |      |         |
|--------------------------------|---------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.32    | 65    | 98786    | 49.10    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 135 | Recovery | =    | 98.20%  |
| 7) Chloroethane-d5             | 1.58    | 69    | 68378    | 49.45    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 130 | Recovery | =    | 98.90%  |
| 11) 1,1-Dichloroethene-d2      | 2.12    | 63    | 151514   | 37.42    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 74.84%  |
| 21) 2-Butanone-d5              | 3.91    | 46    | 162540   | 105.62   | ug/L | 0.00    |
| Spiked Amount                  | 100.000 | Range | 40 - 130 | Recovery | =    | 105.62% |
| 24) Chloroform-d               | 4.38    | 84    | 229080   | 49.57    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 99.14%  |
| 26) 1,2-Dichloroethane-d4      | 5.05    | 65    | 164088   | 51.83    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 103.66% |
| 32) Benzene-d6                 | 5.07    | 84    | 434134   | 51.85    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 103.70% |
| 36) 1,2-Dichloropropane-d6     | 6.09    | 67    | 130691   | 52.00    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 120 | Recovery | =    | 104.00% |
| 41) Toluene-d8                 | 7.34    | 98    | 403781   | 51.09    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 102.18% |
| 43) trans-1,3-Dichloropropene- | 7.64    | 79    | 72394    | 49.75    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 99.50%  |
| 47) 2-Hexanone-d5              | 8.11    | 63    | 120550   | 104.61   | ug/L | 0.00    |
| Spiked Amount                  | 100.000 | Range | 45 - 130 | Recovery | =    | 104.61% |
| 57) 1,1,2,2-Tetrachloroethane- | 10.24   | 84    | 184807   | 51.68    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 65 - 120 | Recovery | =    | 103.36% |
| 64) 1,2-Dichlorobenzene-d4     | 11.65   | 152   | 174652   | 54.87    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 109.74% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072120\  
Data File : VV017567.D  
Acq On : 21 Jul 2020 16:33  
Operator : SY/MD  
Sample : VV0721WBL01  
Misc : 5.0mL/MSVOA V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK64

Quant Time: Jul 22 03:35:38 2020  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVLM072120WMA.M  
Quant Title : VOC Analysis  
QLast Update : Wed Jul 22 03:23:09 2020  
Response via : Initial Calibration

