

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072220\
 Data File : VV017594.D
 Acq On : 22 Jul 2020 18:41
 Operator : SY/MD
 Sample : L3386-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jul 23 06:37:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM072120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 23 05:23:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	298783	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	291699	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	145369	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80297	44.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.08%
7) Chloroethane-d5	1.58	69	64157	51.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.38%
11) 1,1-Dichloroethene-d2	2.12	63	125204	34.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.24%
21) 2-Butanone-d5	3.91	46	137818	98.81	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.81%
24) Chloroform-d	4.37	84	203567	48.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.22%
26) 1,2-Dichloroethane-d4	5.05	65	150238	52.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.72%
32) Benzene-d6	5.07	84	372001	47.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.52%
36) 1,2-Dichloropropane-d6	6.09	67	114766	49.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.16%
41) Toluene-d8	7.33	98	348333	47.38	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.76%
43) trans-1,3-Dichloropropene-	7.64	79	65375	48.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.60%
47) 2-Hexanone-d5	8.11	63	100566	93.81	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.81%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	157851	47.45	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.90%
64) 1,2-Dichlorobenzene-d4	11.65	152	153597	50.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.90%

Target Compounds					Ovalue
16) Methylene chloride	2.52	84	1833	0.813 ug/L	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072220\
Data File : VV017594.D
Acq On : 22 Jul 2020 18:41
Operator : SY/MD
Sample : L3386-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jul 23 06:37:00 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM072120WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 23 05:23:23 2020
Response via : Initial Calibration

