

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072221\
 Data File : VV021698.D
 Acq On : 22 Jul 2021 12:27
 Operator : SY/MD
 Sample : M3122-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Jul 23 01:17:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 23 01:17:03 2021
 Response via : Initial Calibration


Compound	R.T. QIon		Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	475088	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	434880	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	218132	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	169593	50.461	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 100.920%		
7) Chloroethane-d5	1.558	69	136054	58.507	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 117.020%		
11) 1,1-Dichloroethene-d2	2.101	63	227935	40.699	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 81.400%		
21) 2-Butanone-d5	3.863	46	247503	95.544	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 95.540%		
24) Chloroform-d	4.342	84	323405	50.675	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 101.360%		
26) 1,2-Dichloroethane-d4	5.027	65	202510	53.574	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 107.140%		
32) Benzene-d6	5.043	84	677537	52.935	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 105.860%		
36) 1,2-Dichloropropane-d6	6.066	67	212705	53.150	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 106.300%		
41) Toluene-d8	7.313	98	617731	52.238	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 104.480%		
43) trans-1,3-Dichloroprop...	7.619	79	101025	50.640	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 101.280%		
47) 2-Hexanone-d5	8.085	63	154865	85.684	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 85.680%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	244928	46.333	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 92.660%		
66) 1,2-Dichlorobenzene-d4	11.625	152	238434	52.874	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 105.740%		
Target Compounds						Qvalue
13) Acetone	2.156	43	5192	3.160	ug/L	94
25) Chloroform	4.371	83	15025	2.579	ug/L	98
44) trans-1,3-Dichloropropene	7.619	75	4219	0.858	ug/L #	79
48) 2-Hexanone	8.140	43	15815	4.334	ug/L #	84
61) 1,2,3-Trichloropropane	11.249	75	23665	5.849	ug/L #	66


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072221\
Data File : VV021698.D
Acq On : 22 Jul 2021 12:27
Operator : SY/MD
Sample : M3122-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Jul 23 01:17:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 23 01:17:03 2021
Response via : Initial Calibration









#13
Acetone
Concen: 3.160 ug/L
RT: 2.156 min Scan# 347
Delta R.T. -0.000 min
Lab File: VV021698.D
Acq: 22 Jul 2021 12:27




Instrument :
MSVOA_V
ClientSampleId :


Tgt Ion: 43 Resp: 5192
Ion Ratio Lower Upper
43 100
58 35.7 0.0 64.6



#25
Chloroform
Concen: 2.579 ug/L
RT: 4.371 min Scan# 1036
Delta R.T. -0.000 min
Lab File: VV021698.D
Acq: 22 Jul 2021 12:27

Tgt Ion: 83 Resp: 15025
Ion Ratio Lower Upper
83 100
85 64.4 43.8 81.4









#44
trans-1,3-Dichloropropene
Concen: 0.858 ug/L
RT: 7.619 min Scan# 2046
Delta R.T. -0.029 min
Lab File: VV021698.D
Acq: 22 Jul 2021 12:27




Instrument :
MSVOA_V
ClientSampled :


Tgt Ion: 75 Resp: 4219
Ion Ratio Lower Upper
75 100
77 42.9 21.8 40.4#



#48
2-Hexanone
Concen: 4.334 ug/L
RT: 8.140 min Scan# 2208
Delta R.T. 0.003 min
Lab File: VV021698.D
Acq: 22 Jul 2021 12:27


Tgt Ion: 43 Resp: 15815
Ion Ratio Lower Upper
43 100
58 47.7 46.0 69.0
57 10.5 15.6 23.4#
100 4.7 11.0 16.4#





#61
 1,2,3-Trichloropropane
 Concen: 5.849 ug/L
 RT: 11.249 min Scan# 3175
 Delta R.T. 0.974 min
 Lab File: VV021698.D
 Acq: 22 Jul 2021 12:27

Instrument :
 MSVOA_V
 ClientSampleId :



Tgt Ion:	75	Resp:	23665
Ion Ratio	Lower	Upper	
75	100		
77	32.7	26.1	39.1
110	2.5	32.9	49.3#

