

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072318\
 Data File : VV006702.D
 Acq On : 23 Jul 2018 17:56
 Operator : SY/MD
 Sample : VV0723WBL02
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK70

Quant Time: Jul 24 03:17:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 24 02:34:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	246637	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	226499	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	96759	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	65875	5.13	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.60%
7) Chloroethane-d5	1.58	69	52615	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	2.13	63	87959	3.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.80%
20) 2-Butanone-d5	3.96	46	228150	56.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.76%
24) Chloroform-d	4.41	84	126029	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.60%
26) 1,2-Dichloroethane-d4	5.09	65	64572	5.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.00%
32) Benzene-d6	5.10	84	251839	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.20%
36) 1,2-Dichloropropane-d6	6.12	67	83822	5.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.60%
41) Toluene-d8	7.36	98	205297	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
43) trans-1,3-Dichloropropene-	7.67	79	25828	5.03	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.60%
46) 2-Hexanone-d5	8.14	63	180916	55.67	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.34%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	66842	5.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	78734	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072318\
Data File : VV006702.D
Acq On : 23 Jul 2018 17:56
Operator : SY/MD
Sample : VV0723WBL02
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK70

Quant Time: Jul 24 03:17:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 24 02:34:01 2018
Response via : Initial Calibration

