

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072319\
 Data File : VV011962.D
 Acq On : 23 Jul 2019 13:56
 Operator : SY/MD
 Sample : VV0723WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK53

Quant Time: Jul 24 06:03:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 24 05:52:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	140579	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	135180	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	67120	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	35071	4.80	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.00%
7) Chloroethane-d5	1.58	69	28659	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	2.13	63	58344	4.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.20%
20) 2-Butanone-d5	3.96	46	105998	49.63	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.26%
24) Chloroform-d	4.40	84	90036	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
26) 1,2-Dichloroethane-d4	5.08	65	50878	5.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.00%
32) Benzene-d6	5.10	84	186259	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
36) 1,2-Dichloropropane-d6	6.12	67	52496	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.36	98	168425	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
43) trans-1,3-Dichloropropene-	7.66	79	20140	4.60	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.00%
46) 2-Hexanone-d5	8.14	63	85318	48.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.04%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	35547	4.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	67538	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072319\
Data File : VV011962.D
Acq On : 23 Jul 2019 13:56
Operator : SY/MD
Sample : VV0723WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK53

Quant Time: Jul 24 06:03:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 24 05:52:52 2019
Response via : Initial Calibration

