

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072319\
 Data File : VV011994.D
 Acq On : 24 Jul 2019 04:32
 Operator : SY/MD
 Sample : K3960-02
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB9A7

Quant Time: Jul 24 08:53:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 24 07:47:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	148905	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	146275	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	76651	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	33352	4.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.20%
7) Chloroethane-d5	1.59	69	28185	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.60%
11) 1,1-Dichloroethene-d2	2.13	63	57025	3.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.00%
20) 2-Butanone-d5	3.96	46	109386	48.35	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.70%
24) Chloroform-d	4.40	84	94452	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
26) 1,2-Dichloroethane-d4	5.08	65	52345	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
32) Benzene-d6	5.10	84	185041	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
36) 1,2-Dichloropropane-d6	6.12	67	51744	4.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.60%
41) Toluene-d8	7.36	98	174547	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
43) trans-1,3-Dichloropropene-	7.66	79	19652	4.15	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.00%
46) 2-Hexanone-d5	8.14	63	91597	48.14	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.28%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	37450	4.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	73015	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
40) 4-Methyl-2-pentanone	7.28	43	26572	4.597	ug/L	98
42) Toluene	7.43	91	75766	1.640	ug/L	99
52) Ethylbenzene	9.05	91	1068808	20.275	ug/L	99
53) m,p-xylene	9.18	106	1790839	89.245	ug/L	99
54) o-xylene	9.59	106	460106	24.182	ug/L	97
56) Isopropylbenzene	9.97	105	16273	0.306	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072319\
Data File : VV011994.D
Acq On : 24 Jul 2019 04:32
Operator : SY/MD
Sample : K3960-02
Misc : 25ML/MSVOA V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB9A7

Quant Time: Jul 24 08:53:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 24 07:47:25 2019
Response via : Initial Calibration

