

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072320\
 Data File : VV017621.D
 Acq On : 23 Jul 2020 23:03
 Operator : SY/MD
 Sample : L3398-03
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4C51

Quant Time: Jul 24 03:08:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 24 02:33:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	160327	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	158727	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	78492	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	37844	4.94	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.80%
7) Chloroethane-d5	1.58	69	32257	5.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.20%
11) 1,1-Dichloroethene-d2	2.12	63	63285	3.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.80%
20) 2-Butanone-d5	3.93	46	91476	49.95	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.90%
24) Chloroform-d	4.38	84	95215	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	5.06	65	50117	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.80%
32) Benzene-d6	5.07	84	171262	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.10	67	48488	4.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.60%
41) Toluene-d8	7.34	98	154962	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
45) trans-1,3-Dichloropropene-	7.65	79	19249	4.37	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.40%
48) 2-Hexanone-d5	8.11	63	59857	43.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.64%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	32876	4.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	67359	5.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.60%

Target Compounds					Ovalue
34) Trichloroethene	5.94	95	35431	2.823 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072320\
Data File : VV017621.D
Acq On : 23 Jul 2020 23:03
Operator : SY/MD
Sample : L3398-03
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4C51

Quant Time: Jul 24 03:08:13 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 24 02:33:05 2020
Response via : Initial Calibration

