

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072320\
 Data File : VV017622.D
 Acq On : 23 Jul 2020 23:27
 Operator : SY/MD
 Sample : L3398-04
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4C52

Quant Time: Jul 24 03:10:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 24 02:33:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	157481	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	156105	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	72670	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	40731	5.42	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.40%
7) Chloroethane-d5	1.58	69	33195	5.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.40%
11) 1,1-Dichloroethene-d2	2.12	63	63950	3.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.80%
20) 2-Butanone-d5	3.93	46	94387	52.47	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.94%
24) Chloroform-d	4.38	84	95935	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.40%
26) 1,2-Dichloroethane-d4	5.06	65	53279	5.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.60%
32) Benzene-d6	5.07	84	181480	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.40%
36) 1,2-Dichloropropane-d6	6.09	67	52345	5.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.40%
41) Toluene-d8	7.34	98	164768	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
45) trans-1,3-Dichloropropene-	7.65	79	20166	4.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.00%
48) 2-Hexanone-d5	8.11	63	62305	46.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.76%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	34203	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	65095	5.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072320\
Data File : VV017622.D
Acq On : 23 Jul 2020 23:27
Operator : SY/MD
Sample : L3398-04
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4C52

Quant Time: Jul 24 03:10:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 24 02:33:05 2020
Response via : Initial Calibration

