

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072320\
 Data File : VV017624.D
 Acq On : 24 Jul 2020 00:15
 Operator : SY/MD
 Sample : L3398-06
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4C54

Quant Time: Jul 24 03:18:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 24 02:33:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	158673	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	157068	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	73305	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	37096	4.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.00%
7) Chloroethane-d5	1.58	69	30905	5.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.60%
11) 1,1-Dichloroethene-d2	2.12	63	61967	3.65	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.00%
20) 2-Butanone-d5	3.94	46	93408	51.53	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.06%
24) Chloroform-d	4.38	84	94842	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
26) 1,2-Dichloroethane-d4	5.06	65	50745	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.40%
32) Benzene-d6	5.07	84	171875	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
36) 1,2-Dichloropropane-d6	6.09	67	47805	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.34	98	154664	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
45) trans-1,3-Dichloropropene-	7.64	79	19438	4.46	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.20%
48) 2-Hexanone-d5	8.11	63	58226	43.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.16%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	33222	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	62114	5.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072320\
Data File : VV017624.D
Acq On : 24 Jul 2020 00:15
Operator : SY/MD
Sample : L3398-06
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4C54

Quant Time: Jul 24 03:18:46 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 24 02:33:05 2020
Response via : Initial Calibration

