

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072320\
 Data File : VV017626.D
 Acq On : 24 Jul 2020 01:03
 Operator : SY/MD
 Sample : L3398-08
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4C60

Quant Time: Jul 24 03:23:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 24 02:33:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	156051	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	153371	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	72821	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	35062	4.71	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.20%
7) Chloroethane-d5	1.58	69	28862	5.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.20%
11) 1,1-Dichloroethene-d2	2.12	63	60312	3.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.20%
20) 2-Butanone-d5	3.95	46	87324	48.99	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.98%
24) Chloroform-d	4.37	84	91211	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
26) 1,2-Dichloroethane-d4	5.06	65	49430	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.20%
32) Benzene-d6	5.07	84	164972	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
36) 1,2-Dichloropropane-d6	6.10	67	46231	4.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.40%
41) Toluene-d8	7.34	98	149149	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
45) trans-1,3-Dichloropropene-	7.65	79	19326	4.54	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.80%
48) 2-Hexanone-d5	8.12	63	57446	43.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.04%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	31479	4.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	59577	5.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.20%

Target Compounds					Ovalue
34) Trichloroethene	5.94	95	6319	0.521 ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072320\
Data File : VV017626.D
Acq On : 24 Jul 2020 01:03
Operator : SY/MD
Sample : L3398-08
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4C60

Quant Time: Jul 24 03:23:25 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 24 02:33:05 2020
Response via : Initial Calibration

