

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072320\
 Data File : VV017605.D
 Acq On : 23 Jul 2020 15:43
 Operator : SY/MD
 Sample : VV0723WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK53

Quant Time: Jul 24 02:01:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 24 01:49:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	185523	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	182552	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	89287	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	41645	4.70	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.00%
7) Chloroethane-d5	1.58	69	33255	4.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.20%
11) 1,1-Dichloroethene-d2	2.12	63	67554	3.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.00%
20) 2-Butanone-d5	3.96	46	101829	48.05	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.10%
24) Chloroform-d	4.38	84	102892	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.06	65	56617	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	5.07	84	195390	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.10	67	54664	4.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.80%
41) Toluene-d8	7.34	98	175746	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
45) trans-1,3-Dichloropropene-	7.65	79	23398	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
48) 2-Hexanone-d5	8.12	63	70773	45.05	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.10%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	36460	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	69130	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072320\
Data File : VV017605.D
Acq On : 23 Jul 2020 15:43
Operator : SY/MD
Sample : VV0723WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK53

Quant Time: Jul 24 02:01:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 24 01:49:19 2020
Response via : Initial Calibration

