

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072419\
 Data File : VV012002.D
 Acq On : 24 Jul 2019 11:07
 Operator : SY/MD
 Sample : VV0724WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK55

Quant Time: Jul 25 04:28:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 25 04:12:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	138367	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	130381	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	62732	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	32346	4.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.00%
7) Chloroethane-d5	1.59	69	28864	5.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.00%
11) 1,1-Dichloroethene-d2	2.13	63	55349	3.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.40%
20) 2-Butanone-d5	3.97	46	115424	54.91	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.82%
24) Chloroform-d	4.40	84	93246	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.08	65	53698	5.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.60%
32) Benzene-d6	5.10	84	190028	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.60%
36) 1,2-Dichloropropane-d6	6.12	67	54322	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.40%
41) Toluene-d8	7.36	98	170238	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
43) trans-1,3-Dichloropropene-	7.66	79	21050	4.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.80%
46) 2-Hexanone-d5	8.14	63	92873	54.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.52%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	38733	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	67787	5.61	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072419\
Data File : VV012002.D
Acq On : 24 Jul 2019 11:07
Operator : SY/MD
Sample : VV0724WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK55

Quant Time: Jul 25 04:28:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 25 04:12:02 2019
Response via : Initial Calibration

