

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072420\
 Data File : VV017662.D
 Acq On : 24 Jul 2020 21:48
 Operator : SY/MD
 Sample : L3446-05
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4C63

Quant Time: Jul 25 03:13:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 25 00:05:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	150545	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	148103	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	71826	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	33227	4.62	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.40%
7) Chloroethane-d5	1.58	69	27576	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.20%
11) 1,1-Dichloroethene-d2	2.13	63	77386	4.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.00%
20) 2-Butanone-d5	3.92	46	72579	42.20	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.40%
24) Chloroform-d	4.38	84	85419	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	5.06	65	44446	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
32) Benzene-d6	5.07	84	148837	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
36) 1,2-Dichloropropane-d6	6.09	67	41737	4.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.00%
41) Toluene-d8	7.34	98	134591	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
45) trans-1,3-Dichloropropene-	7.65	79	16115	3.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.40%
48) 2-Hexanone-d5	8.11	63	48022	37.68	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.36%
59) 1,1,2,2-Tetrachloroethane-	10.23	84	26509	4.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	55437	5.10	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.00%

Target Compounds

					Ovalue
10) 1,1,2-Trichloro-1,2,2-trif	2.13	101	2631	0.293 ug/L	92
12) 1,1-Dichloroethene	2.14	96	34460	4.251 ug/L	91
19) 1,1-Dichloroethane	3.21	63	2749	0.148 ug/L	98
34) Trichloroethene	5.94	95	30755	2.626 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072420\
Data File : VV017662.D
Acq On : 24 Jul 2020 21:48
Operator : SY/MD
Sample : L3446-05
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampled :
F4C63

Quant Time: Jul 25 03:13:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 25 00:05:26 2020
Response via : Initial Calibration

