

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072523\
 Data File : VW031744.D
 Acq On : 25 Jul 2023 16:27
 Operator : SY/MD
 Sample : 03712-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH2W6

Quant Time: Jul 26 02:57:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 26 02:25:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	262424	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	300631	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	136568	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	116067	43.386	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	86.780%		
7) Chloroethane-d5	1.532	69	111369	45.426	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	90.860%		
11) 1,1-Dichloroethene-d2	2.059	65	53939	43.558	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	87.120%		
21) 2-Butanone-d5	3.793	46	210485	113.704	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	113.700%		
24) Chloroform-d	4.256	84	275434	50.077	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.160%		
26) 1,2-Dichloroethane-d4	4.947	65	173360	50.657	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	101.320%		
32) Benzene-d6	4.966	84	412008	43.279	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.560%		
36) 1,2-Dichloropropane-d6	5.995	67	161260	45.096	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	90.200%		
41) Toluene-d8	7.249	98	309891	38.836	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	77.680%#		
43) trans-1,3-Dichloroprop...	7.558	79	62939	40.948	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	81.900%		
47) 2-Hexanone-d5	8.030	63	132560	98.853	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	98.850%		
56) 1,1,2,2-Tetrachloroeth...	10.159	84	245535	49.834	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	99.660%		
66) 1,2-Dichlorobenzene-d4	11.567	152	140820	50.928	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	101.860%		
Target Compounds					Qvalue	
13) Acetone	2.117	43	11177	6.308	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072523\
Data File : VV031744.D
Acq On : 25 Jul 2023 16:27
Operator : SY/MD
Sample : 03712-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH2W6

Quant Time: Jul 26 02:57:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 26 02:25:02 2023
Response via : Initial Calibration

