

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072524\
 Data File : VV036685.D
 Acq On : 25 Jul 2024 00:58
 Operator : SY/MD
 Sample : PP3335-10DL200X
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Jul 25 04:30:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 25 03:06:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	327478	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	343246	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	146008	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	100829	41.841	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	83.680%	
7) Chloroethane-d5	1.526	69	91162	44.182	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	88.360%	
11) 1,1-Dichloroethene-d2	2.053	65	48726	42.272	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	84.540%	
21) 2-Butanone-d5	3.780	46	109492	88.653	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	88.650%	
24) Chloroform-d	4.243	84	220002	44.589	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.180%	
26) 1,2-Dichloroethane-d4	4.934	65	142144	48.377	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.760%	
32) Benzene-d6	4.953	84	393900	43.750	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.500%	
36) 1,2-Dichloropropane-d6	5.982	67	119329	46.687	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	93.380%	
41) Toluene-d8	7.239	98	351465	40.020	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	80.040%	
43) trans-1,3-Dichloroprop...	7.551	79	51357	37.125	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	74.260%	
47) 2-Hexanone-d5	8.021	63	77821	82.268	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	82.270%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	168840	42.631	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	85.260%	
66) 1,2-Dichlorobenzene-d4	11.558	152	155385	51.627	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	103.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072524\
Data File : VV036685.D
Acq On : 25 Jul 2024 00:58
Operator : SY/MD
Sample : PP3335-10DL200X
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Jul 25 04:30:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072424WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 25 03:06:08 2024
Response via : Initial Calibration

