

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072524\
 Data File : VW036690.D
 Acq On : 25 Jul 2024 02:58
 Operator : SY/MD
 Sample : P3335-17 200X
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E20N0

Quant Time: Jul 25 04:31:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 25 03:06:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	337256	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	357939	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	155168	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	102232	41.193	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.380%
7) Chloroethane-d5	1.529	69	91073	42.860	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.720%
11) 1,1-Dichloroethene-d2	2.056	65	49423	41.633	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.260%
21) 2-Butanone-d5	3.780	46	103855	81.651	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.650%
24) Chloroform-d	4.243	84	221533	43.597	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.200%
26) 1,2-Dichloroethane-d4	4.937	65	139062	45.956	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.920%
32) Benzene-d6	4.953	84	406536	43.300	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.600%
36) 1,2-Dichloropropane-d6	5.985	67	118732	44.547	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.100%
41) Toluene-d8	7.239	98	361919	39.519	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.040%#
43) trans-1,3-Dichloroprop...	7.551	79	51881	35.965	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.920%
47) 2-Hexanone-d5	8.021	63	75490	76.528	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.530%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	167778	40.624	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.240%
66) 1,2-Dichlorobenzene-d4	11.558	152	161408	50.462	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.920%
Target Compounds						
					Qvalue	
34) Trichloroethene	5.838	95	6315	1.918	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072524\
Data File : VV036690.D
Acq On : 25 Jul 2024 02:58
Operator : SY/MD
Sample : P3335-17 200X
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E20N0

Quant Time: Jul 25 04:31:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072424WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 25 03:06:08 2024
Response via : Initial Calibration

