

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072618\
 Data File : VV006709.D
 Acq On : 26 Jul 2018 17:29
 Operator : SY/MD
 Sample : J4148-03
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0SW8

Quant Time: Jul 27 01:14:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 27 00:52:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	240733	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	225825	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	93501	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	92156	7.36	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	147.20%#
7) Chloroethane-d5	1.58	69	64671	6.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	131.40%#
11) 1,1-Dichloroethene-d2	2.13	63	122277	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.40%
20) 2-Butanone-d5	3.97	46	206472	52.74	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.48%
24) Chloroform-d	4.41	84	145242	6.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	123.60%
26) 1,2-Dichloroethane-d4	5.09	65	73346	6.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	128.00%
32) Benzene-d6	5.10	84	313927	6.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	135.20%#
36) 1,2-Dichloropropane-d6	6.12	67	98488	6.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	125.60%
41) Toluene-d8	7.37	98	272358	6.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	134.20%#
43) trans-1,3-Dichloropropene-	7.67	79	30901	6.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	120.80%
46) 2-Hexanone-d5	8.14	63	164446	50.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.52%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	67931	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	95294	6.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	136.80%#

Target Compounds					Ovalue
13) Acetone	2.21	43	5007	2.01 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072618\
Data File : VV006709.D
Acq On : 26 Jul 2018 17:29
Operator : SY/MD
Sample : J4148-03
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0SW8

Quant Time: Jul 27 01:14:36 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 27 00:52:23 2018
Response via : Initial Calibration

