

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072622\
 Data File : VW026962.D
 Acq On : 26 Jul 2022 19:36
 Operator : SY/MD
 Sample : N3843-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUE2

Quant Time: Jul 27 01:43:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 27 01:39:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	499744	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	501996	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	207711	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	178037	42.477	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	84.960%	
7) Chloroethane-d5	1.561	69	155129	45.750	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	91.500%	
11) 1,1-Dichloroethene-d2	2.101	63	246845	32.735	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	65.480%	
21) 2-Butanone-d5	3.883	46	286300	101.338	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	101.340%	
24) Chloroform-d	4.342	84	385526	46.736	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.480%	
26) 1,2-Dichloroethane-d4	5.027	65	233485	49.483	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.960%	
32) Benzene-d6	5.047	84	742489	48.931	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.860%	
36) 1,2-Dichloropropane-d6	6.066	67	249144	50.949	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	101.900%	
41) Toluene-d8	7.313	98	590734	44.493	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	88.980%	
43) trans-1,3-Dichloroprop...	7.619	79	96502	47.097	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	94.200%	
47) 2-Hexanone-d5	8.088	63	195621	91.874	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	91.870%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	366791	49.375	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	98.740%	
66) 1,2-Dichlorobenzene-d4	11.622	152	226513	58.128	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	116.260%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072622\
Data File : VV026962.D
Acq On : 26 Jul 2022 19:36
Operator : SY/MD
Sample : N3843-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUE2

Quant Time: Jul 27 01:43:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 27 01:39:23 2022
Response via : Initial Calibration

