

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072622\
 Data File : VW026985.D
 Acq On : 27 Jul 2022 05:23
 Operator : SY/MD
 Sample : N3880-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUG3

Quant Time: Jul 27 07:04:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 27 02:43:35 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	515988	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	507134	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	218762	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	227639	52.601	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	105.200%	
7) Chloroethane-d5	1.564	69	192396	54.954	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	109.900%	
11) 1,1-Dichloroethene-d2	2.105	63	310042	39.822	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	79.640%	
21) 2-Butanone-d5	3.883	46	301996	103.529	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	103.530%	
24) Chloroform-d	4.342	84	440956	51.773	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	103.540%	
26) 1,2-Dichloroethane-d4	5.027	65	264465	54.284	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	108.560%	
32) Benzene-d6	5.047	84	885493	57.764	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	115.520%	
36) 1,2-Dichloropropane-d6	6.066	67	284858	57.662	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	115.320%	
41) Toluene-d8	7.313	98	729605	54.396	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	108.800%	
43) trans-1,3-Dichloroprop...	7.622	79	92253	44.567	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	89.140%	
47) 2-Hexanone-d5	8.088	63	223063	103.701	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	103.700%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	389997	51.967	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	103.940%	
66) 1,2-Dichlorobenzene-d4	11.622	152	263845	64.288	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	128.580%#	
Target Compounds						
20) cis-1,2-Dichloroethene	3.905	96	40270	9.064	ug/L	93
34) Trichloroethene	5.912	95	75859	18.101	ug/L	98
46) Tetrachloroethene	7.982	164	15246794	4663.715	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072622\
Data File : VV026985.D
Acq On : 27 Jul 2022 05:23
Operator : SY/MD
Sample : N3880-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUG3

Quant Time: Jul 27 07:04:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 27 02:43:35 2022
Response via : Initial Calibration

