

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072622\
 Data File : VW026949.D
 Acq On : 26 Jul 2022 13:51
 Operator : SY/MD
 Sample : VW0726WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK231

Quant Time: Jul 27 01:40:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 27 01:39:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	506374	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	515664	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	223147	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	200888	47.301	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.600%
7) Chloroethane-d5	1.564	69	172778	50.287	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.580%
11) 1,1-Dichloroethene-d2	2.101	63	277919	36.374	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.740%
21) 2-Butanone-d5	3.886	46	286978	100.248	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.250%
24) Chloroform-d	4.339	84	404811	48.432	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.860%
26) 1,2-Dichloroethane-d4	5.024	65	248305	51.935	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.860%
32) Benzene-d6	5.043	84	802399	51.477	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.960%
36) 1,2-Dichloropropane-d6	6.063	67	261083	51.975	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.940%
41) Toluene-d8	7.313	98	650902	47.726	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.460%
43) trans-1,3-Dichloroprop...	7.619	79	105051	49.910	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.820%
47) 2-Hexanone-d5	8.088	63	196593	89.883	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.880%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	376422	49.328	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.660%
66) 1,2-Dichlorobenzene-d4	11.622	152	249479	59.593	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	119.180%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072622\
Data File : VV026949.D
Acq On : 26 Jul 2022 13:51
Operator : SY/MD
Sample : VV0726WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK231

Quant Time: Jul 27 01:40:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 27 01:39:23 2022
Response via : Initial Calibration

