

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072623\
 Data File : VW031759.D
 Acq On : 26 Jul 2023 11:18
 Operator : SY/MD
 Sample : 03754-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH2Y3

Quant Time: Jul 27 05:19:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 27 05:18:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	246884	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	280350	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	126357	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	152822	60.720	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 121.440%			
7) Chloroethane-d5	1.532	69	143938	62.406	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 124.820%			
11) 1,1-Dichloroethene-d2	2.060	65	68035	58.399	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 116.800%			
21) 2-Butanone-d5	3.789	46	217900	125.119	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 125.120%			
24) Chloroform-d	4.252	84	319013	61.651	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 123.300%			
26) 1,2-Dichloroethane-d4	4.947	65	207183	64.351	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 128.700%#			
32) Benzene-d6	4.966	84	510866	57.546	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 115.100%			
36) 1,2-Dichloropropane-d6	5.995	67	204053	61.190	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 122.380%#			
41) Toluene-d8	7.249	98	396562	53.293	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 106.580%			
43) trans-1,3-Dichloroprop...	7.558	79	78633	54.859	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 109.720%			
47) 2-Hexanone-d5	8.030	63	132496	105.953	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 105.950%			
56) 1,1,2,2-Tetrachloroeth...	10.156	84	258273	56.211	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 112.420%			
66) 1,2-Dichlorobenzene-d4	11.567	152	176092	68.830	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 137.660%#			
Target Compounds					Qvalue	
13) Acetone	2.114	43	37318	22.387	ug/L	96
25) Chloroform	4.285	83	11365	2.190	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072623\
Data File : VV031759.D
Acq On : 26 Jul 2023 11:18
Operator : SY/MD
Sample : 03754-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH2Y3

Quant Time: Jul 27 05:19:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 27 05:18:30 2023
Response via : Initial Calibration

