

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072623\
 Data File : VW031761.D
 Acq On : 26 Jul 2023 12:05
 Operator : SY/MD
 Sample : 03754-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH2X5

Quant Time: Jul 27 05:40:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 27 05:18:30 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.542	114	245213	50.000	ug/L	0.00
28) Chlorobenzene-d5		8.792	117	268656	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.191	152	120103	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.278	65	156101	62.446	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	124.900%	
7) Chloroethane-d5		1.532	69	142259	62.099	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	124.200%	
11) 1,1-Dichloroethene-d2		2.063	65	67096	57.986	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	115.980%	
21) 2-Butanone-d5		3.789	46	219184	126.714	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	126.710%	
24) Chloroform-d		4.256	84	325543	63.342	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	126.680%#	
26) 1,2-Dichloroethane-d4		4.947	65	212640	66.496	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	133.000%#	
32) Benzene-d6		4.966	84	522685	61.440	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	122.880%	
36) 1,2-Dichloropropane-d6		5.992	67	202852	63.478	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	126.960%#	
41) Toluene-d8		7.249	98	393839	55.231	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.460%	
43) trans-1,3-Dichloroprop...		7.558	79	82128	59.792	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	119.580%	
47) 2-Hexanone-d5		8.030	63	132960	110.952	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.950%	
56) 1,1,2,2-Tetrachloroeth...		10.159	84	287564	65.311	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	130.620%#	
66) 1,2-Dichlorobenzene-d4		11.567	152	180281	74.137	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	148.280%#	
Target Compounds							Qvalue
13) Acetone		2.114	43	64229	38.793	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072623\
Data File : VV031761.D
Acq On : 26 Jul 2023 12:05
Operator : SY/MD
Sample : 03754-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH2X5

Quant Time: Jul 27 05:40:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 27 05:18:30 2023
Response via : Initial Calibration

