

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072623\
 Data File : VV031756.D
 Acq On : 26 Jul 2023 10:02
 Operator : SY/MD
 Sample : VV0726WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK374

Quant Time: Jul 27 05:19:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 27 05:18:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	327489	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	353585	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	160084	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	153698	46.038	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	92.080%	
7) Chloroethane-d5	1.532	69	149061	48.721	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	97.440%	
11) 1,1-Dichloroethene-d2	2.060	65	70572	45.667	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	91.340%	
21) 2-Butanone-d5	3.786	46	218703	94.671	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	94.670%	
24) Chloroform-d	4.249	84	314492	45.818	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.640%	
26) 1,2-Dichloroethane-d4	4.944	65	198688	46.523	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.040%	
32) Benzene-d6	4.963	84	504810	45.086	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.180%	
36) 1,2-Dichloropropane-d6	5.992	67	196893	46.814	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	93.620%	
41) Toluene-d8	7.249	98	381328	40.632	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	81.260%	
43) trans-1,3-Dichloroprop...	7.554	79	79473	43.961	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	87.920%	
47) 2-Hexanone-d5	8.027	63	126333	80.100	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	80.100%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	260927	45.027	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	90.060%	
66) 1,2-Dichlorobenzene-d4	11.564	152	172828	53.322	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	106.640%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072623\
Data File : VV031756.D
Acq On : 26 Jul 2023 10:02
Operator : SY/MD
Sample : VV0726WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK374

Quant Time: Jul 27 05:19:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072523WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 27 05:18:30 2023
Response via : Initial Calibration

