

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV072720\
 Data File : VV017682.D
 Acq On : 27 Jul 2020 17:06
 Operator : SY/MD
 Sample : L3346-07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4C34

Quant Time: Jul 28 06:32:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 28 04:52:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	158193	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	157491	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	75316	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	30011	3.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.40%
7) Chloroethane-d5	1.58	69	25718	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.00%
11) 1,1-Dichloroethene-d2	2.12	63	52035	3.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.40%
20) 2-Butanone-d5	3.93	46	78120	43.23	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.46%
24) Chloroform-d	4.38	84	83287	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
26) 1,2-Dichloroethane-d4	5.06	65	44247	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
32) Benzene-d6	5.07	84	147376	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.00%
36) 1,2-Dichloropropane-d6	6.09	67	39824	4.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.60%
41) Toluene-d8	7.34	98	129721	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.00%
45) trans-1,3-Dichloropropene-	7.64	79	17606	4.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.40%
48) 2-Hexanone-d5	8.11	63	50138	36.99	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.98%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	26790	3.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	54929	4.82	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV072720\
Data File : VV017682.D
Acq On : 27 Jul 2020 17:06
Operator : SY/MD
Sample : L3346-07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4C34

Quant Time: Jul 28 06:32:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 28 04:52:47 2020
Response via : Initial Calibration

