

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072722\
 Data File : VW027020.D
 Acq On : 28 Jul 2022 00:20
 Operator : SY/MD
 Sample : N3880-11DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUG5DL

Quant Time: Jul 28 06:17:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 28 02:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	450990	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	444380	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	187241	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	198980	52.606	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 105.220%			
7) Chloroethane-d5	1.564	69	169318	55.332	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 110.660%			
11) 1,1-Dichloroethene-d2	2.101	63	265035	38.947	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 77.900%			
21) 2-Butanone-d5	3.895	46	202518	79.432	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 79.430%			
24) Chloroform-d	4.342	84	363603	48.844	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 97.680%			
26) 1,2-Dichloroethane-d4	5.027	65	217658	51.115	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 102.240%			
32) Benzene-d6	5.043	84	739697	55.067	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 110.140%			
36) 1,2-Dichloropropane-d6	6.066	67	238082	54.999	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 110.000%			
41) Toluene-d8	7.313	98	588342	50.059	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 100.120%			
43) trans-1,3-Dichloroprop...	7.619	79	88678	48.890	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 97.780%			
47) 2-Hexanone-d5	8.091	63	181079	96.071	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 96.070%			
56) 1,1,2,2-Tetrachloroeth...	10.213	84	333766	50.754	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 101.500%			
66) 1,2-Dichlorobenzene-d4	11.622	152	211422	60.187	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 120.380%#			
Target Compounds						
					Qvalue	
10) 1,1,2-Trichloro-1,2,2-...	2.101	101	2561	0.686	ug/L #	26
20) cis-1,2-Dichloroethene	3.908	96	23249	5.987	ug/L	98
34) Trichloroethene	5.915	95	25207	6.864	ug/L	99
46) Tetrachloroethene	7.972	164	160105	55.889	ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072722\
Data File : VV027020.D
Acq On : 28 Jul 2022 00:20
Operator : SY/MD
Sample : N3880-11DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUG5DL

Quant Time: Jul 28 06:17:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 28 02:51:09 2022
Response via : Initial Calibration









#10
 1,1,2-Trichloro-1,2,2-trifluoroethane
 Concen: 0.686 ug/L
 RT: 2.101 min Scan# 31
 Delta R.T. -0.010 min
 Lab File: VV027020.D
 Acq: 28 Jul 2022 00:20




Instrument :
 MSVOA_V
 ClientSampleId :
 DBUG5DL


Tgt Ion:101 Resp: 2561
 Ion Ratio Lower Upper
 101 100
 85 0.0 33.1 49.7#
 151 5.7 57.1 85.7#



#20
 cis-1,2-Dichloroethene
 Concen: 5.987 ug/L
 RT: 3.908 min Scan# 892
 Delta R.T. 0.003 min
 Lab File: VV027020.D
 Acq: 28 Jul 2022 00:20

Tgt Ion: 96 Resp: 23249
 Ion Ratio Lower Upper
 96 100
 61 116.3 83.0 154.1
 68 0.0 0.0 0.0





#34

Trichloroethene

Concen: 6.864 ug/L

RT: 5.915 min Scan# 1

Delta R.T. 0.003 min

Lab File: VV027020.D

Acq: 28 Jul 2022 00:20

Instrument :

MSVOA_V

ClientSampleId :

DBUG5DL

Tgt Ion: 95 Resp: 25207





Ion Ratio Lower Upper

95 100

97 66.7 45.9 85.3

132 99.1 68.1 126.5

130 100.1 70.9 131.7



#46

Tetrachloroethene

Concen: 55.889 ug/L

RT: 7.972 min Scan# 2156

Delta R.T. -0.000 min

Lab File: VV027020.D

Acq: 28 Jul 2022 00:20

Tgt Ion: 164 Resp: 160105

Ion Ratio Lower Upper

164 100

129 97.3 66.3 123.1

131 96.4 63.1 117.3

166 128.7 86.2 160.2

