

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072722\
 Data File : VW027008.D
 Acq On : 27 Jul 2022 19:01
 Operator : SY/MD
 Sample : N3880-04DL 200X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUF7DL

Quant Time: Jul 28 01:45:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 28 01:41:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	509074	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	503468	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	219598	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	147897	34.639	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	69.280%	
7) Chloroethane-d5	1.565	69	147938	42.829	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	85.660%	
11) 1,1-Dichloroethene-d2	2.101	63	228439	29.739	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	59.480%#	
21) 2-Butanone-d5	3.883	46	316728	110.054	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	110.050%	
24) Chloroform-d	4.343	84	390582	46.482	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.960%	
26) 1,2-Dichloroethane-d4	5.027	65	234539	48.795	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.600%	
32) Benzene-d6	5.047	84	750988	49.346	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.700%	
36) 1,2-Dichloropropane-d6	6.066	67	254545	51.901	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	103.800%	
41) Toluene-d8	7.313	98	607096	45.592	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	91.180%	
43) trans-1,3-Dichloroprop...	7.619	79	95853	46.643	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	93.280%	
47) 2-Hexanone-d5	8.088	63	232131	108.702	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	108.700%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	373688	50.156	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	100.320%	
66) 1,2-Dichlorobenzene-d4	11.622	152	237273	57.593	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	115.180%	
Target Compounds						Qvalue
46) Tetrachloroethene	7.973	164	105887	32.625	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072722\
Data File : VV027008.D
Acq On : 27 Jul 2022 19:01
Operator : SY/MD
Sample : N3880-04DL 200X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUF7DL

Quant Time: Jul 28 01:45:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 28 01:41:31 2022
Response via : Initial Calibration

