

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072722\
 Data File : VW027018.D
 Acq On : 27 Jul 2022 23:31
 Operator : SY/MD
 Sample : N3843-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUP3

Quant Time: Jul 28 06:17:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 28 02:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	473024	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	465969	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	204972	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	212525	53.569	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 107.140%	
7) Chloroethane-d5	1.564	69	174247	54.291	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 108.580%	
11) 1,1-Dichloroethene-d2	2.101	63	276068	38.679	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 77.360%	
21) 2-Butanone-d5	3.889	46	203230	75.998	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 76.000%	
24) Chloroform-d	4.342	84	372517	47.710	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 95.420%	
26) 1,2-Dichloroethane-d4	5.027	65	224635	50.296	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 100.600%	
32) Benzene-d6	5.043	84	769740	54.649	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 109.300%	
36) 1,2-Dichloropropane-d6	6.066	67	243664	53.681	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 107.360%	
41) Toluene-d8	7.313	98	613251	49.761	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 99.520%	
43) trans-1,3-Dichloroprop...	7.619	79	91396	48.054	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 96.100%	
47) 2-Hexanone-d5	8.088	63	191873	97.081	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 97.080%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	345117	50.049	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 100.100%	
66) 1,2-Dichlorobenzene-d4	11.622	152	227394	59.134	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 118.260%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072722\
Data File : VV027018.D
Acq On : 27 Jul 2022 23:31
Operator : SY/MD
Sample : N3843-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUP3

Quant Time: Jul 28 06:17:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 28 02:51:09 2022
Response via : Initial Calibration

