

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072722\
 Data File : VW027030.D
 Acq On : 28 Jul 2022 04:25
 Operator : SY/MD
 Sample : N3897-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUQ7

Quant Time: Jul 28 06:19:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 28 02:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	450968	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	445286	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	188839	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	200096	52.903	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.800%
7) Chloroethane-d5	1.565	69	167450	54.725	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.440%
11) 1,1-Dichloroethene-d2	2.102	63	275998	40.560	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.120%
21) 2-Butanone-d5	3.892	46	183082	71.812	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	71.810%
24) Chloroform-d	4.343	84	367635	49.388	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.780%
26) 1,2-Dichloroethane-d4	5.027	65	222382	52.227	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.460%
32) Benzene-d6	5.044	84	744344	55.300	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.600%
36) 1,2-Dichloropropane-d6	6.066	67	236903	54.615	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.240%
41) Toluene-d8	7.314	98	602981	51.200	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.400%
43) trans-1,3-Dichloroprop...	7.619	79	87668	48.234	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.460%
47) 2-Hexanone-d5	8.088	63	172713	91.446	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.450%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	335655	50.938	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.880%
66) 1,2-Dichlorobenzene-d4	11.622	152	222049	62.677	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	125.360%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072722\
Data File : VV027030.D
Acq On : 28 Jul 2022 04:25
Operator : SY/MD
Sample : N3897-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUQ7

Quant Time: Jul 28 06:19:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 28 02:51:09 2022
Response via : Initial Calibration

