

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072722\
 Data File : VW026996.D
 Acq On : 27 Jul 2022 14:03
 Operator : SY/MD
 Sample : VW0727WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK234

Quant Time: Jul 28 01:42:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 28 01:41:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	461646	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	464259	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	201734	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	238332	61.555	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 123.100%		
7) Chloroethane-d5	1.564	69	189382	60.461	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 120.920%		
11) 1,1-Dichloroethene-d2	2.101	63	314999	45.221	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 90.440%		
21) 2-Butanone-d5	3.886	46	250518	95.991	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 95.990%		
24) Chloroform-d	4.339	84	403660	52.973	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 105.940%		
26) 1,2-Dichloroethane-d4	5.024	65	238034	54.610	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 109.220%		
32) Benzene-d6	5.043	84	823764	58.700	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 117.400%		
36) 1,2-Dichloropropane-d6	6.066	67	257106	56.851	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 113.700%		
41) Toluene-d8	7.313	98	685214	55.805	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 111.600%		
43) trans-1,3-Dichloroprop...	7.619	79	105031	55.426	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 110.860%		
47) 2-Hexanone-d5	8.088	63	170820	86.747	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 86.750%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	349335	50.847	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 101.700%		
66) 1,2-Dichlorobenzene-d4	11.622	152	243179	64.254	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 128.500%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072722\
Data File : VV026996.D
Acq On : 27 Jul 2022 14:03
Operator : SY/MD
Sample : VV0727WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK234

Quant Time: Jul 28 01:42:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM07222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 28 01:41:31 2022
Response via : Initial Calibration

