

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072722\
 Data File : VW0727033.D
 Acq On : 28 Jul 2022 05:38
 Operator : SY/MD
 Sample : N3899-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUR5

Quant Time: Jul 28 06:20:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 28 02:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	429746	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	436077	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	183307	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	196870	54.621	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 109.240%	
7) Chloroethane-d5	1.565	69	165450	56.741	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 113.480%	
11) 1,1-Dichloroethene-d2	2.101	63	269132	41.504	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 83.000%	
21) 2-Butanone-d5	3.889	46	165724	68.214	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 68.210%	
24) Chloroform-d	4.343	84	360981	50.889	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 101.780%	
26) 1,2-Dichloroethane-d4	5.027	65	211039	52.011	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 104.020%	
32) Benzene-d6	5.047	84	729971	55.378	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 110.760%	
36) 1,2-Dichloropropane-d6	6.066	67	230865	54.347	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 108.700%	
41) Toluene-d8	7.314	98	591867	51.317	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 102.640%	
43) trans-1,3-Dichloroprop...	7.619	79	84402	47.418	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 94.840%	
47) 2-Hexanone-d5	8.088	63	162734	87.982	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 87.980%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	325321	50.412	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 100.820%	
66) 1,2-Dichlorobenzene-d4	11.622	152	213601	62.112	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 124.220%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072722\
Data File : VV027033.D
Acq On : 28 Jul 2022 05:38
Operator : SY/MD
Sample : N3899-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUR5

Quant Time: Jul 28 06:20:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 28 02:51:09 2022
Response via : Initial Calibration

