

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072821\
 Data File : VV021704.D
 Acq On : 28 Jul 2021 12:54
 Operator : SY/MD
 Sample : VV0728MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK221

Quant Time: Jul 29 02:03:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 29 02:02:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	473136	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	441143	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	215196	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	139202	41.589	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.180%
7) Chloroethane-d5	1.561	69	114009	49.229	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.460%
11) 1,1-Dichloroethene-d2	2.105	63	187324	33.586	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.180%
21) 2-Butanone-d5	3.873	46	231759	89.836	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.840%
24) Chloroform-d	4.346	84	277628	43.682	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.360%
26) 1,2-Dichloroethane-d4	5.031	65	171178	45.472	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.940%
32) Benzene-d6	5.050	84	580488	44.709	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.420%
36) 1,2-Dichloropropane-d6	6.069	67	183371	45.170	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.340%
41) Toluene-d8	7.317	98	532358	44.379	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.760%
43) trans-1,3-Dichloroprop...	7.622	79	86893	42.938	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.880%
47) 2-Hexanone-d5	8.085	63	150363	82.012	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.010%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	226484	42.236	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.480%
66) 1,2-Dichlorobenzene-d4	11.625	152	205524	46.198	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072821\
Data File : VV021704.D
Acq On : 28 Jul 2021 12:54
Operator : SY/MD
Sample : VV0728MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK221

Quant Time: Jul 29 02:03:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM071921WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 29 02:02:57 2021
Response via : Initial Calibration

