

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072822\
 Data File : VW027064.D
 Acq On : 29 Jul 2022 03:21
 Operator : SY/MD
 Sample : N3790-11ME
 Misc : 5.73g/5.0mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUK5ME

Quant Time: Jul 29 05:16:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 29 02:10:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	611477	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	577110	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	239629	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	222074	38.174	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	76.340%	
7) Chloroethane-d5	1.561	69	132455	28.737	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	57.480%#	
11) 1,1-Dichloroethene-d2	2.082	63	260511	26.304	ug/L	-0.02
Spiked Amount	50.000	Range 60 - 125	Recovery	=	52.600%#	
21) 2-Butanone-d5	3.915	46	211396	63.729	ug/L	0.03
Spiked Amount	100.000	Range 40 - 130	Recovery	=	63.730%	
24) Chloroform-d	4.339	84	408841	39.021	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	78.040%	
26) 1,2-Dichloroethane-d4	5.024	65	245391	41.258	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	82.520%	
32) Benzene-d6	5.040	84	853400	42.110	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.220%	
36) 1,2-Dichloropropane-d6	6.066	67	269937	43.518	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	87.040%	
41) Toluene-d8	7.314	98	691790	37.766	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	75.540%#	
43) trans-1,3-Dichloroprop...	7.622	79	102680	37.474	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	74.940%	
47) 2-Hexanone-d5	8.108	63	182845	68.570	ug/L	0.02
Spiked Amount	100.000	Range 45 - 130	Recovery	=	68.570%	
56) 1,1,2,2-Tetrachloroeth...	10.220	84	350755	37.504	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	75.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	235976	43.806	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.620%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072822\
Data File : VV027064.D
Acq On : 29 Jul 2022 03:21
Operator : SY/MD
Sample : N3790-11ME
Misc : 5.73g/5.0mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUK5ME

Quant Time: Jul 29 05:16:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 29 02:10:37 2022
Response via : Initial Calibration

