

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072822\
 Data File : VV027038.D
 Acq On : 28 Jul 2022 12:20
 Operator : SY/MD
 Sample : VV0728MBL01
 Misc : 5.00g/5.0mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK236

Quant Time: Jul 29 02:01:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 29 01:53:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	572289	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	553000	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	252758	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	179615	32.990	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.980%
7) Chloroethane-d5	1.561	69	120775	27.997	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	56.000%#
11) 1,1-Dichloroethene-d2	2.092	63	240727	25.971	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	51.940%#
21) 2-Butanone-d5	3.889	46	310081	99.880	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.880%
24) Chloroform-d	4.336	84	375677	38.311	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.620%
26) 1,2-Dichloroethane-d4	5.021	65	228146	40.985	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.980%
32) Benzene-d6	5.037	84	743124	38.267	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.540%
36) 1,2-Dichloropropane-d6	6.059	67	247789	41.689	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.380%
41) Toluene-d8	7.310	98	589052	33.559	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	67.120%#
43) trans-1,3-Dichloroprop...	7.616	79	101627	38.707	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.420%
47) 2-Hexanone-d5	8.092	63	214624	83.997	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.000%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	389829	43.500	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	257981	45.404	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.800%
Target Compounds						Qvalue
71) Naphthalene	13.506	128	26531	1.715	ug/L	99
72) 1,2,3-Trichlorobenzene	13.744	180	3861	0.801	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072822\
Data File : VV027038.D
Acq On : 28 Jul 2022 12:20
Operator : SY/MD
Sample : VV0728MBL01
Misc : 5.00g/5.0mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK236

Quant Time: Jul 29 02:01:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 29 01:53:08 2022
Response via : Initial Calibration

