

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072822\
 Data File : VV027039.D
 Acq On : 28 Jul 2022 14:37
 Operator : SY/MD
 Sample : VV0728WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK237

Quant Time: Jul 29 02:01:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 29 01:53:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	464679	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	472553	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	205613	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	172436	39.005	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	78.020%		
7) Chloroethane-d5	1.564	69	152946	43.665	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	87.340%		
11) 1,1-Dichloroethene-d2	2.101	63	237075	31.500	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	63.000%		
21) 2-Butanone-d5	3.883	46	215910	85.652	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	85.650%		
24) Chloroform-d	4.342	84	390316	49.022	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.040%		
26) 1,2-Dichloroethane-d4	5.027	65	226030	50.008	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.020%		
32) Benzene-d6	5.043	84	714218	43.040	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.080%		
36) 1,2-Dichloropropane-d6	6.066	67	246992	48.630	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	97.260%		
41) Toluene-d8	7.313	98	565045	37.672	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	75.340%#		
43) trans-1,3-Dichloroprop...	7.619	79	98835	44.052	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	88.100%		
47) 2-Hexanone-d5	8.088	63	209635	96.012	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	96.010%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	395567	51.654	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	103.300%		
66) 1,2-Dichlorobenzene-d4	11.622	152	244271	52.848	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	105.700%		
Target Compounds						
39) cis-1,3-Dichloropropene	6.982	75	5393	0.986	ug/L #	74
61) 1,2,3-Trichloropropane	11.246	75	23242	5.390	ug/L #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072822\
Data File : VV027039.D
Acq On : 28 Jul 2022 14:37
Operator : SY/MD
Sample : VV0728WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK237

Quant Time: Jul 29 02:01:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 29 01:53:08 2022
Response via : Initial Calibration

